TPOO2 HS HIGH SENSITIVITY TRANSVERSE HALL PROBE DATA SHEET

HIRST



VERSION 1.0 - 2012

Hirst Magnetic Instruments Ltd.

Tesla House, Tregoniggie, Falmouth Cornwall, TR11 4SN, United Kingdom Tel +44 (0) 1326 372734 Fax +44 (0) 1326 378069 Email <u>sales@hirst-magnetics.com</u> VAT Reg. No. GB730 0277 74 Reg. in England No. 1564814 Websites <u>http://www.hirst-magnetics.com/</u> <u>http://www.gaussmeter.co.uk/</u> The Hirst Magnetic Instruments Ltd TP002-HS is a high sensitivity Hall probe for use with Hirst Magnetic Instruments Ltd. GM07 and GM08 Gaussmeters.



HIRST

The TP002-HS is an active semiconductor Hall element which incorporates a focusing magnetic circuit. This magnetic circuit focuses

the magnetic flux experienced by the probe in the Hall element to offer a significant boost in sensitivity.

Although the ferrite magnetic circuit does exhibit some hysteresis over the operating range, this effect is quite small, less than typically $\pm 0.2\%$.

If at any time during operation or storage, the probe is exposed to high magnetic fields, greatly in excess of the operating range of the probe, a hysteresis error may occur, creating an offset of reading, normally not greater than $\pm 100 \ \mu$ T.

In the event of this happening the Null function (Utilities menu) can be used to remove the offset. Alternatively, demagnetise the Hall probe can help. This procedure is normally unnecessary.

TP002-HS compatibility with GM07 and GM08 models



The TP002-HS can be used immediately with Hirst Magnetic Instruments Ltd GM07 and GM08 Gaussmeter serial numbers: GM07-0924 and GM08-0691 and above

If you wish to use a GM07 or GM08 with an earlier serial number, please contact Hirst Magnetics at <u>sales@hirst-magnetics.com</u> to check on compatibility. (Please use TP002HS compatibility) and include your GM07 or GM08 serial number.

Calibration

Due to the construction of the Hall element in the TP002-HS, it is not suitable for calibration nor calibration checks in conventional reference magnets. The TP002-HS must be calibrated in a uniform magnetic field, not in a reference magnet.

It is recommended that the TP002-HS and GM07 or GM08 Gaussmeters are returned annually for calibration. Please contact our sales office on <u>sales@hirst-magnetics.com</u>.

Operating ranges

HIRST

When operating with the GM07 and GM08, the TP002-HS automatically enables 3 ranges of operation.

- 0 ±29.99 mT (milli Tesla) 0 - ±2.999 mT (milli Tesla)
- $0 \pm 299.9 \,\mu\text{T}$ (micro Tesla)
- 0 ±299.9 Gauss
- 0 ±29.99 Gauss
- 0 ±2.999 Gauss
- 0 ±299.9 Oersted
- 0 ±29,99 Oersted
- 0 ±2.999 Oersted



TRANSVERSE PROBE TP002 HS



Please note that in the interests of continual improvement, specifications may be changed without notice. Please see website for current information.

Tel +44 (0) 1326 372734 Fax +44 (0) 1326 378069 Email <u>sales@hirst-magnetics.com</u> VAT Reg. No. GB730 0277 74 Reg. in England No. 1564814 Websites <u>http://www.hirst-magnetics.com/</u> <u>http://www.gaussmeter.co.uk/</u>